

GLASSWARE CLEANING PROCEDURE (CLIA-88)
UPDATED: AUGUST 2011

CPT CODE: N/A

All glassware is:

- Cleaned with Alconox® detergent
- Rinsed 3X with sterile water
- Air dried

Reviewed By: S. Perlman, D. Fuerbringer

Scott B. Perlman, MD, MS
Chief, Nuclear Medicine

Derek Fuerbringer, CNMT
Manager, Nuclear Medicine